



Food & Beverage Series

GLUCOSE MEASUREMENT IN CANNED GREEN BEANS

Application Note 210LS
YSI Life Sciences



Life Sciences
Data for Life.™

a xylem brand

Glucose Measurement in Canned Green Beans

YSI Life Sciences Application Note 210LS

INTRODUCTION

Dextrose (D-glucose) concentrations in complex matrices such as canned green beans can be measured directly and quickly using the YSI 2900 Series Biochemistry Analyzer. YSI's unique enzyme technology provides for specific glucose measurement. Measurements are virtually unaffected by color, turbidity, density, pH, or the presence of reducing substances.

When a sample is injected into the sample chamber, the glucose diffuses into the membrane containing glucose oxidase. The glucose is immediately oxidized to hydrogen peroxide and D-glucono- δ -lactone. The hydrogen peroxide is detected amperometrically at the platinum electrode surface. The current flow at the electrode is directly proportional to the hydrogen peroxide concentration, and hence to glucose concentration.

I. MATERIALS & SETUP

- A. YSI 2900 Series Biochemistry Analyzer - equipped with a 2365 Glucose Membrane and 2357 Buffer.
- B. Glucose standards (2.50 g/L, 9.00 g/L).
- C. Buffer Diluent (40 g/L NaH_2PO_4 , 10 g/L Na_2HPO_4 in reagent water).
- D. Connect the 2900 Series instrument to a suitable power source.
- E. Perform the instrument and membrane daily checks described in the Operations Manual..
- F. Volumetric glassware (Class A recommended).
- G. The following instrument setup is recommended:
Sample Size 25 μL

Probe A Parameters

Chemistry	Glucose
Unit	g/L
Calibrator	2.50 g/L
End Point	30 Sec

Autocal Parameters

Temperature	1°C
Time	30 Min
Sample	5 Sam
Cal Shift	2%

II. METHOD

- A. Empty contents of can into a clean dry blender.
- B. Blend until the sample is liquid, about ten minutes.
- C. Transfer 25 to 50 grams of the blended sample to a 100 mL volumetric flask, using buffer diluent to rinse and dilute. Fill the flask to the mark with buffer and mix. Allow the solution to equilibrate for about twenty minutes before analysis.
- D. Calibrate the 2900 Series instrument with a 2.50 g/L glucose standard solution.
- E. Check the linearity of the membrane at least once a day by injection of a glucose linearity check solution (9.00 g/L). Refer to the Operators Manual for specifications.
- F. Assay the sample prepared in C by aspiration into the 2900 Series. The linear range of the system is 0.05 to 9.00 g/L glucose. If the value reported exceeds this, further dilution is required.*
- G. Calibrate frequently as described in the Operations Manual.

* The linearity of glucose on the 2900 series instrument can be increased to 0.05 to 25.0 g/L. This can be done by decreasing the sample size to 10 μL and checking the linearity with a 25.0 g/L solution.

III. CALCULATIONS

To calculate % glucose, multiply the reported value by the appropriate dilution factor.

continued

Example: 50.02 grams of blended green beans were diluted to 100 mL in a Class A volumetric flask. When assayed, the value reported was 2.93 g/L glucose.

$\frac{\% \text{ Glucose: } 2.93 \text{ g/L} \times 0.100\text{L}}{50.02 \text{ g}}$	$= 0.0059 \text{ gram glucose/gram green beans}$ $= 0.59\% \text{ (w/w)}$
--	---

ORDERING INFORMATION

YSI Part Numbers:

- 2900 Biochemistry Analyzer
- 2365 Glucose Membrane Kit
- 2776 Glucose Standard Solution (2.50 g/L)
- 1531 Glucose Standard Solution (9.00 g/L)
- 2777 Glucose Standard Solution (25.0 g/L)
- 2357 Buffer Kit
- 2363 Potassium Ferrocyanide Test Solution
- 2392 NaCl Solution (for membrane installation)

For further information, please contact:

YSI Life Sciences

1725 Brannum Lane | Yellow Springs, Ohio 45387

Website: ysi.com | Email: support@ysi.com

Telephone: (937) 767-7241 | Fax: (937) 767-9320



Life Sciences
Data for Life.™

a xylem brand

YSI Life Sciences develops and manufactures scientific instruments, sensors and systems that serve a variety of scientific and industrial markets worldwide. YSI has a long history in the life sciences and bioanalytical markets, most notably with our introduction of the world's first commercial whole blood glucose analyzer in 1975. Today there are over 10,000 YSI instruments installed around the world, trusted in critical situations to provide the most accurate data in the shortest time.

ysi.com/lifesciences

YSI Life Sciences | 1725 Brannum Lane | Yellow Springs, Ohio 45387

Tel +1.800.659.8895 | +1.937.767.7241 | support@ysi.com

© 2015 Xylem, Inc. All rights reserved.

xylem
Let's Solve Water