

Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241•800 765-4974•Fax 937 767-9353

Product Category Membrane
 Product Model Number 2329
 Lot Number 17D100148
 Method Amperometric testing with standards

Time of Testing	Determination	Specification	Results
Initial	Average Calibration Sensitivity	10.0-30.0 nA	19.3 nA
	Average Linearity	14.3-15.7mmol/L	14.63 mmol/L
	Potassium Ferrocyanide	-0.3 - 0.3mmol/L	0.01 mmol/L

Conforms to acceptance criteria

Completed by: Christy Woodford 4/12/2017
 Your Name



Lisa M. Abel
 Director, Quality and Corporate
 Responsibility