

# Certificate of Analysis



**YSI Incorporated**

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category            Membrane  
 Product Model Number    2365  
 Lot Number                 17G100024  
 Method                     Amperometric testing with standards

Time of Testing	Determination	Specification	Results
Initial	Average Calibration Sensitivity	8.0-45.0 nA	09.8 nA
	Average Linearity	8.55-9.45 g/L	8.86 g/L
	Potassium Ferrocyanide	0.00-0.05	0 g/L

Conforms to acceptance criteria

Completed by:             Amanda Mills 07/13/2017  
 \_\_\_\_\_  
 Your Name



Lisa M. Abel  
 Director, Quality and Corporate  
 Responsibility