

Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category	Buffer and solid products
Product Model Number	2363
Lot Number	17H100343
Method	Verification of mass

Determination	Specification	Result
Average Mass	1.15-1.35 grams	1.28 g

Conforms to acceptance criteria

Completed by: Kristin Stover
Your Name

Date: 8/17/2017



A handwritten signature in blue ink, appearing to read 'W. Mabel'.

Director, Quality and Corporate
Responsibility