

# Certificate of Analysis



**YSI Incorporated**

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category Bioanalytical Standard  
Product Model Number 2776  
Lot Number 17H100702  
Method Comparative enzymatic testing of Production standard vs Metrology standard

| Chemistry Tested | Determination   | Tolerance      | Result   |
|------------------|---|----------------|----------|
| GLUCOSE          | (Average of Production)-(Average of Metrology)          | $\pm 0.03$ g/L | 0.01 g/L |
|                  | Coefficient of variation for production standard (CV,%) | $\leq 2.0\%$   | 0.8 %    |
| LACTATE          | (Average of Production)-(Average of Metrology)          | $\pm 0.01$ g/L | 0.00 g/L |
|                  | Coefficient of variation for production standard (CV,%) | $\leq 2.0\%$   | 0.8 %    |

Conforms to acceptance criteria

Completed by: Kristin Stover

Date: 8/29/2017



A handwritten signature in blue ink, appearing to read 'Lisa M. Abel'.

Lisa M. Abel  
Director, Quality and Corporate  
Responsibility