

# Certificate of Analysis



**YSI Incorporated**

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category Bioanalytical Standard  
Product Model Number 2716  
Lot Number 18M100244  
Method Comparative enzymatic testing of Production standard vs Metrology standard

| Chemistry Tested | Determination   | Tolerance | Result  |
|------------------|---|-----------|---------|
| GLUCOSE          | (Average of Production)-(Average of Metrology)          | ±35mg/L   | 19 mg/L |
|                  | Coefficient of variation for production standard (CV,%) | ≤2.0%     | 0.6 %   |

Conforms to acceptance criteria

Completed by: Kristin Stover

Date: 12/13/2018



A handwritten signature in blue ink, appearing to read 'Lisa M. Abel'.

Lisa M. Abel  
Director, Quality and Corporate  
Responsibility