

Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category Buffer and solid products
Product Model Number 2363
Lot Number 19B100775
Method Verification of mass

Determination	Specification	Result
Average Mass	1.15-1.35 grams	1.28 g

Conforms to acceptance criteria

Completed by: Kristin Stover
Your Name

Date: 2/27/2019



A handwritten signature in blue ink, appearing to read 'W. Mabel'.

Director, Quality and Corporate
Responsibility