Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241•800 765-4974•Fax 937 767-9353

a xylem brand

Product Category	Bioanalytical Standard
Product Model Number	2785
Lot Number	<u>19E100318</u>
Method	Comparative enzymatic testing of Production standard vs. Metrology standard

Chemistry Tested	Determination	Tolerance	Result
Glucose	(Average of production) - (average of metrology)	±0.13g/L	0.04g/L
Glucose	Coefficient of variation for production standard (CV, %)	≤2%	1.3%
Galactose	(Average of production) - (average of metrology)	±0.13g/L	0.09g/L
Galactose	Coefficient of variation for production standard (CV, %)	≤2%	1.1%

Conforms to acceptance criteria

Completed by: Heather Brooks

Date: 5/14/19

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.



Dregory W. Popp

Gregory W. Popp Quality Manager, Analytics