Certificate of Analysis



Product Category

Lot Number

Product Model Number

YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241•800 765-4974•Fax 937 767-9353

2784

19F100133

Bioanalytical Standard

standard (CV, %)

ethod	Comparative enzymatic testing of Production standard vs. Metrology standard			
Chemistry Tested	Determination	Tolerance	Result	
LACTOSE	(Average of production) - (average of metrology)	±0.3 g/L	0.02 g/L	
	Coefficient of variation for production standard (CV, %)	≤2%	0.48%	
	(Average of production) - (average of metrology)	N/A	N/A	
	Coefficient of variation for production	N/A	N/A	

Conforms to acce	ptance criteria		
Completed by:	Lucas Stover	Date:	06/12/2019



Lisa M. Abel Director, Quality and Corporate Responsibility

Turalel