Certificate of Analysis

YSI Incorporated
Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA
Phone 937 767-7241•800 765-4974•Fax 937 767-9353

Product Category: Bioanalytical Standard

Product Model Number: 2785

Lot Number: 19F100375

Method: Comparative enzymatic testing of Production standard vs. Metrology standard

Chemistry Tested	Determination	Tolerance	Result
Galactose	(Average of production) - (average of metrology)	± 0.13 g/L	0.01g/L
	Coefficient of variation for production standard (CV, %)	≤ 2 %	1.0%
Glucose	(Average of production) - (average of metrology)	± 0.13 g/L	0.00g/L
	Coefficient of variation for production standard (CV, %)	≤ 2 %	1.2%

Conforms to acceptance criteria

Completed by: Lucas Stover Date: 06/26/2019

Lisa M. Abel
Director, Quality and Corporate Responsibility