Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241•800 765-4974•Fax 937 767-9353

Product Category Membrane

Product Model Number 2329

Lot Number 19F100606

Method Amperometric testing with standards

Time of Testing	Determination	Specification	Results
Initial	Average Calibration Sensitivity	6.0-30.0 nA	13.0 nA
	Average Linearity	14.3-15.7mmol/L	14.84 mmol/L
	Potassium Ferrocyanide	-0.3 - 0.3mmol/L	0.01 mmol/L

Conforms to acceptance criteria

Completed by: Sandy Crawford 6/25/2019

Your Name

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.

STATE OF THE PARTY OF THE PARTY

Lisa M. Abel
Director, Quality and Corporate
Responsibility