Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241•800 765-4974•Fax 937 767-9353

Product Category	Bioanalytical Standard	
Product Model Number	2785	
Lot Number	19H100802	
Method	Comparative enzymatic testing	of Production standard vs. Metrology standard

Chemistry Tested	Determination	Tolerance	Result
Galactose	(Average of production) - (average of metrology)	± 0.13 g/L	0.10g/L
	Coefficient of variation for production standard (CV, %)	≤ 2 %	0.90%
Glucose	(Average of production) - (average of metrology)	± 0.13 g/L	0.02g/L
	Coefficient of variation for production standard (CV, %)	≤ 2 %	1.04%

Conforms to acceptance criteria

Completed by:

David Stoler

Date:

9-4-2019



mallel

Lisa M. Abel Director, Quality and Corporate Responsibility