

# Certificate of Analysis



**YSI Incorporated**  
Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA  
Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category                      Bioanalytical Standard  
Product Model Number                2726  
Lot Number                                19J100133  
Method                                      Comparative enzymatic testing of Production standard vs. Metrology standard

Chemistry Tested	Determination	Tolerance	Result
Methanol Std A 1.0 g/L	(Average of production) - (average of metrology)	$\pm 20$ mg/L	12.3 mg/L
	Coefficient of variation for production standard (CV, %)	$\leq 2$ %	1.79%
Methanol Std B 2.50 g/L	(Average of production) - (average of metrology)	$\pm 50$ mg/L	27 mg/L
	Coefficient of variation for production standard (CV, %)	$\leq 2$ %	0.49%

Conforms to acceptance criteria

Completed by: David Stoler

Date: 9-11-2019



Lisa M. Abel  
Director, Quality and Corporate  
Responsibility