

# Certificate of Analysis



**YSI Incorporated**

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category             Membrane  
Product Model Number   2329  
Lot Number                 19J100696  
Method                     Amperometric testing with standards

Time of Testing	Determination	Specification	Results
Initial	Average Calibration Sensitivity	6.0-30.0 nA	12.7 nA
	Average Linearity	14.3-15.7mmol/L	14.99 mmol/L
	Potassium Ferrocyanide	-0.3 - 0.3mmol/L	0.01 mmol/L

Conforms to acceptance criteria

Completed by:                             Christy Woodford                     10/1/2019  
   Your Name

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.



Lisa M. Abel  
Director, Quality and Corporate  
Responsibility