Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241•800 765-4974•Fax 937 767-9353

Product Category	Bioanalytical Standard	
Product Model Number	2784	_
Lot Number	19K100448	_
Method	Comparative enzymatic testing	ng of Production standard vs. Metrology standard

Chemistry Tested	Determination	Tolerance	Result
LACTOSE	(Average of production) - (average of metrology)	±0.3 g/L	0.04 g/L
	Coefficient of variation for production standard (CV, %)	≤2%	0.8%
	(Average of production) - (average of metrology)	N/A	N/A
	Coefficient of variation for production standard (CV, %)	N/A	N/A

Conforms to acce	ptance criteria			_
Completed by:	Lucas Stover	Date:	10/23/2019	



Lisa M. Abel Director, Quality and Corporate Responsibility

Tomalel