Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category	Bioanalytical Standard		
Product Model Number	2783		
Lot Number	19L100210		
Method	Comparative enzymatic testing of Production standard vs. Metrology standard		

Chemistry Tested	Determination	Tolerance	Result
Lactose	(Average of production) - (average of metrology)	± 0.1 g/L	0.00g/L
	Coefficient of variation for production standard (CV, %)	≤ 2 %	1.33%
	(Average of production) - (average of metrology)	N/A	N/A
	Coefficient of variation for production standard (CV, %)	N/A	N/A

Conforms to acce	Conforms to acceptance criteria				
Completed by:	David Stoler	Date:	11-14-2019		



Lisa M. Abel Director, Quality and Corporate Responsibility

Thralel