

# Certificate of Analysis



**YSI Incorporated**  
 Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA  
 Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category Bioanalytical Standard

Product Model Number 2792

Lot Number 19L100217

Method Comparative enzymatic testing of Production standard vs. Metrology standard

Chemistry Tested	Determination	Tolerance	Result
ETHANOL STD A 1.0G/L	(Average of production) - (average of metrology)	±20mg/L	5.05 mg/L
	Coefficient of variation for production standard (CV, %)	≤2%	0.95%
ETHANOL STD B 0.5G/L	(Average of production) - (average of metrology)	±10mg/L	6.75 mg/L
	Coefficient of variation for production standard (CV, %)	≤2%	0.61%

Conforms to acceptance criteria

Completed by: David Stoler Date: 11-14-2019



Lisa M. Abel  
 Director, Quality and Corporate  
 Responsibility