Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241•800 765-4974•Fax 937 767-9353

Product Category	Bioanalytical Standard	
Product Model Number	2783	
Lot Number	20A100861	
Method	Comparative enzymatic testing	g of Production standard vs. Metrology standard

Chemistry Tested	Determination	Tolerance	Result
Lactose	(Average of production) - (average of metrology)	± 0.1 g/L	0.03g/L
	Coefficient of variation for production standard (CV, %)	$\leq 2 \%$	1.2%
	(Average of production) - (average of metrology)	N/A	N/A
	Coefficient of variation for production standard (CV, %)	N/A	N/A

Conforms to acceptance criteria

Completed by:

Lucas Stover

Date:

02/04/2020



mullel

Lisa M. Abel Director, Quality and Corporate Responsibility