

# Certificate of Analysis



**YSI Incorporated**

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

---

Product Category	Membrane
Product Model Number	2365
Lot Number	21C100552
Method	Amperometric testing with standards

---

Time of Testing	Determination	Specification	Results
Initial	Average Calibration Sensitivity	8.0-45.0 nA	09.2 nA
	Average Linearity	8.55-9.45 g/L	9.10 g/L
	Potassium Ferrocyanide	0.00-0.05	g/L

---

Conforms to acceptance criteria

Completed by: Sandy Crawford  
Your Name

3/24/2021



A handwritten signature in blue ink that reads 'Gregory W. Popp'.

Gregory W. Popp  
Quality Manager Analytics