Certificate of Analysis

YSI Incorporated
Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA
Phone 937 767-7241•800 765-4974•Fax 937 767-9353

Product Category	Membrane
Product Model Number	2754
Lot Number	21F100555
Method	Amperometric testing with standards

Time of Testing	Determination	Specification	Results
Initial	Average Calibration Sensitivity	8.0-40.0 nA	09.0 nA
	Average Linearity	9.50-10.50mmol/L	9.98 mmol/L
	Potassium Ferrocyanide	< 0.4mmol/L	0.04 mmol/L

Conforms to acceptance criteria

Completed by: Christy Woodford 6/24/2021
Your Name

ISO 9001 CERTIFIED YSI INCORPORATED

Gregory W. Popp

Gregory W. Popp
Quality Manager Analytics