

Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category Membrane
Product Model Number 2735
Lot Number 22A100529
Method Amperometric testing with standards

| Time of Testing | Determination | Specification | Results |
|-----------------|---------------------------------|---------------|-------------|
| Initial | Average Calibration Sensitivity | 8.0-40.0 nA | 10.0 nA |
| | Average Linearity | 7.6-8.4mmol/L | 7.99 mmol/L |
| | Potassium Ferrocyanide | < 0.4mmol/L | 0.01 mmol/L |

Conforms to acceptance criteria

Completed by:

Christy Woodford

1/25/2022

Your Name



A handwritten signature in blue ink that reads 'Gregory W. Popp'.

Gregory W. Popp
Quality Manager Analytics