

# Certificate of Analysis



**YSI Incorporated**

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

---

|                      |                                     |
|----------------------|-------------------------------------|
| Product Category     | Membrane                            |
| Product Model Number | 2365                                |
| Lot Number           | 22J100373                           |
| Method               | Amperometric testing with standards |

---

| Time of Testing | Determination                   | Specification | Results  |
|-----------------|---------------------------------|---------------|----------|
| Initial         | Average Calibration Sensitivity | 8.0-45.0 nA   | 12.2 nA  |
|                 | Average Linearity               | 8.55-9.45 g/L | 8.88 g/L |
|                 | Potassium Ferrocyanide          | 0.00-0.05     | 0.00 g/L |

---

Conforms to acceptance criteria

Completed by: Sandy Crawford  
Your Name

9/23/2022



A handwritten signature in blue ink that reads 'Gregory W. Popp'.

Gregory W. Popp  
Quality Manager Analytics