

# Certificate of Analysis



**YSI Incorporated**

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

---

Product Category	Membrane
Product Model Number	2771
Lot Number	23A100285
Method	Amperometric testing with standards

---

Time of Testing	Determination	Specification	Results
Initial	Average Calibration Sensitivity	>5 nA	15.28nA
	Average Linearity	427 - 473 mg/L	451.00mg/L
	Potassium Ferrocyanide Interference	<15 mg/L	0.01mg/L
Day 2	Average Calibration Sensitivity	>5 nA	14.04nA
	Average Linearity	427 - 473 mg/L	433.20mg/L
	Potassium Ferrocyanide Interference	<15 mg/L	0.01mg/L

---

Conforms to acceptance criteria

Completed by:

Christy Woodford

1/18/2023

---

Your Name



A handwritten signature in blue ink that reads 'Gregory W. Popp'.

Gregory W. Popp  
Quality Manager Analytics