

Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category Membrane
Product Model Number 2725
Lot Number 23C100093
Method Amperometric testing with standards

| Time of Testing | Determination | Specification | Results |
|-----------------|---------------------------------|-----------------|---------|
| Initial | Average Calibration Sensitivity | 5.0-25.0 nA | 6.27 |
| | Average Linearity | 2.375-2.625 g/L | 2.47 |
| | Potassium Ferrocyanide | -0.2-0.2 g/L | 0.00 |
| Final | Average Calibration Sensitivity | 5.0-25.0 nA | 6.75 |
| | Average Linearity | 2.375-2.625 g/L | 2.49 |
| | Potassium Ferrocyanide | -0.2-0.2 g/L | 0.00 |

Conforms to acceptance criteria

Completed by: _____ Christy Woodford _____ 3/13/2023
Your Name



A handwritten signature in blue ink that reads 'Gregory W. Popp'.

Gregory W. Popp
Quality Manager