

Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category Membrane
Product Model Number 2365
Lot Number 23C100461
Method Amperometric testing with standards

Time of Testing	Determination	Specification	Results
Initial	Average Calibration Sensitivity	8.0-45.0 nA	12.2 nA
	Average Linearity	8.55-9.45 g/L	8.94 g/L
	Potassium Ferrocyanide	0.00-0.05	0.00 g/L

Sandy Crawford

3/24/2023

Conforms to acceptance criteria

Completed by:

Your Name



A handwritten signature in blue ink that reads 'Gregory W. Popp'.

Gregory W. Popp
Quality Manager Analytics