

# Certificate of Analysis



**YSI Incorporated**

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category            Membrane  
Product Model Number    2725  
Lot Number                 23C100588  
Method                      Amperometric testing with standards

Time of Testing	Determination	Specification	Results
Initial	Average Calibration Sensitivity	5.0-25.0 nA	6.27
	Average Linearity	2.375-2.625 g/L	
	Potassium Ferrocyanide	-0.2-0.2 g/L	0.00
Final	Average Calibration Sensitivity	5.0-25.0 nA	7.29
	Average Linearity	2.375-2.625 g/L	
	Potassium Ferrocyanide	-0.2-0.2 g/L	0.00

Conforms to acceptance criteria

Completed by: \_\_\_\_\_ Christy Woodford \_\_\_\_\_ 4/11/2023  
Your Name



A handwritten signature in blue ink that reads 'Gregory W. Popp'.

Gregory W. Popp  
Quality Manager