

Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category Bioanalytical Standard
Product Model Number 2792
Lot Number 23D100439
Method Comparative enzymatic testing of Production standard vs Metrology standard

Chemistry Tested	Determination	Tolerance	Result
ETHANOL PART A	(Average of Production)-(Average of Metrology)	±20mg/L	8.95 mg/L
	Coefficient of variation for production standard (CV,%)	≤2.0%	0.33 %
ETHANOL PART B	(Average of Production)-(Average of Metrology)	±10mg/L	0.35 mg/L
	Coefficient of variation for production standard (CV,%)	≤2.0%	1.32 %

Conforms to acceptance criteria

Completed by: Amanda Fenner

Date: 4/26/2023

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.



A handwritten signature in blue ink that reads 'Gregory W. Popp'.

Gregory W. Popp
Quality Manager, Analytics