

Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category Bioanalytical Standard
Product Model Number 2792
Lot Number 23F100514
Method Comparative enzymatic testing of Production standard vs Metrology standard

| Chemistry Tested | Determination | Tolerance | Result |
|------------------|---|-----------|-----------|
| ETHANOL PART A | (Average of Production)-(Average of Metrology) | ±20mg/L | 0.40 mg/L |
| | Coefficient of variation for production standard (CV,%) | ≤2.0% | 0.40 % |
| ETHANOL PART B | (Average of Production)-(Average of Metrology) | ±10mg/L | 1.00 mg/L |
| | Coefficient of variation for production standard (CV,%) | ≤2.0% | 0.56 % |

Conforms to acceptance criteria

Completed by: Amanda Fenner

Date: 6/26/2023

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.



A handwritten signature in blue ink that reads 'Gregory W. Popp'.

Gregory W. Popp
Quality Manager, Analytics